Open Tender Bidding

TENDER DOCUMENT

FOR

PROCUREMENT OF

SCIENTIFIC EQUIPMENT

FOR

DRS-III PROGRAMME
DEPARTMENT OF BIOCHEMISTRY
OSMANIA UNIVERSITY

Coordinator
DRS-III PROGRAMME
DEPARTMENT OF BIOCHEMISTRY
OSMANIA UNIVERSITY
HYDERABAD – 500 007

Newspaper Advertisement

Last date 03-08-2017

Procurement of SCIENTIFIC EQUIPMENT

Coordinator
DRS-III PROGRAMME
DEPARTMENT OF BIOCHEMISTRY
OSMANIA UNIVERSITY
HYDERABAD – 500 007

Interested Bidders may visit www.osmania.ac.in

For details please contact:

Coordinator
DRS-III PROGRAMME
DEPARTMENT OF BIOCHEMISTRY, OU
HYDERABAD – 500 007
040-27682469.27682245

	Notice Inviting Tender Details		
S.No	Description		
1.	Department Name	Biochemistry, Osmania University, Hyderabad	
2	Dua accesio a Face	DRS-III Programme, Dept. of Biochemistry, Osmania University,	
2.	Procuring For	Hyderabad	
3.	Circle/Division	NA	
4.	Tender Number	126/ SAP-DRS-III/Bioc/15	
5.	Tender Subject	Supply, Installation, & Commissioning of Scientific Equipment	
6.	Period Of Contract	THREE YEARS	
7.	Form Of Contract	Schedule Wise	
8.	Tender Type	Open	
9.	Tender Category	Products (both Hardware and Software)	
10.	EMD / Bid Security (INR)	5% of the Order value or maximum of Rs.1,00,000/- (Enclose in separate cover)	
		In the form of Demand Draft in favor of "Coordinator, DRS-III	
11.	EMD/Bid Security Payable To	Programme, Dept. of Biochemistry," OU, Hyderabad."	
12.	Non-refundable Tender Fee	Rs.5,000 /- (Enclose in separate cover) for each equipment	
13.	Tender Fee	The DD should be drawn in the favor of "Coordinator, DRS-III Programme, Dept. of Biochemistry Osmania University, Hyderabad". The Applications can be downloaded for the university website www.osmaina.ac.in . If the application is downloaded from the website, the DD for Rs.5000/- must be drawn in favor of the Coordinator, DRS-III Programme, Dept. of Biochemistry Osmania University, Hyderabad, University College of Science, Osmania University, Hyderabad. This DD can be submitted along with the Tender bids in a separate cover. Note: Please do not enclose the DD inside the bid. It should be in a separate envelope. Application Fee is Non Transferable.	
14.	Schedule Sale opening date	12.07.2017 at 2.00 PM	
15.	Schedule Sale Closing Date	03.08.2017 at 3.00 PM	
16.	Bid Submission Closing Date	03.08.2017 at 3.00 PM	
17.	Bid Submission	In person / Post. However, for postal delay, University cannot be held responsible.	
18.	Pre-Qualification/Technical Bid Opening Date (Qualification and Eligibility Stage)	NA	
	1	3	

19.	Technical Specifications Bid Opening Date(Technical Bid Stage)	03-08-2017 at 4.00 PM
20.	Price Bid Opening Date (Financial Bid Stage)	03-08-2017 at 4.00 PM
21.	Place of Tender Opening	Committee Room, Administrative Building, Osmania University, Hyderabad
22.	Officer Inviting Bids/Contact Person	Coordinator, DRS-III Programme, Dept. of Biochemistry, Osmania University, Hyderabad
23.	Address/E-mail ID	suryasingh.oubioc@gmail.com
24.		Ph:+91-40-27682469, 27682245
	Contact Details/Telephone, Fax	As per the tender document Annexure-02
25.	Eligibility Criterion	•
26.	Procedure For Bid Submission	Tenders shall be submitted exclusively in person /Post. The tenderer can download the tender document from the University website www.osmania.ac.in The tenderers who are desirous of participating shall submit their technical bids, price bids as per the standard formats Annexure-11 and Annexure-12 available in the Tender document. The tenderer should submit the following documents to the Coordinator, DRS-III Programme, Dept. of Biochemistry, Osmania University, Hyderabad. A DD/Pay order drawn in favor of "Coordinator, DRS-III Programme, Dept. of Biochemistry, Osmania University, Hyderabad" for the amount Rs 5,000/- towards Tender fee (Application fee) for each equipment. A DD/Pay order/drawn in favor of "Coordinator, DRS-III Programme, Dept. of Biochemistry, Osmania University, Hyderabad"T.S" for the amount 5% of the Order value or maximum of Rs.1,00,000/- (EMD fee) (Enclose in separate cover) Technical Bid (Annexure-11) along with supporting documents. Commercial Bid (Annexure-12) All the supporting documents as mentioned in the eligibility criteria (all the points). Documents in support of Technical Specifications along with make and models of all the items as per the list mentioned in Annexure-05 duly mentioning the make. Supporting documents of standard certifications. Annexure-02 of tender document duly signed with office seal as a token of acceptance of our standard terms and conditions. Latest income tax clearance certificate. List of customers, to whom the bidder had supplied identical materials in the past along with P.O details and performance report. Annexure-08 of bidding document. The Coordinator, DRS-III Programme, Dept. of Biochemistry, Hyderabad, will not hold any risk and responsibility for nonvisibility of the scanned document or the loss in transit.
27.	General Terms and Conditions	As per tender documents.

Contents of the Tender Document

Ι.	Schedule of Quantity	Annexure – 01
2.	Eligibility Criteria & Special Terms and conditions	Annexure – 02
3.	List of Documents to be enclosed	Annexure – 03
4.	List of Addresses	Annexure – 04
5.	Technical Specifications	Annexure – 05
6.	Instructions to Bidders	Annexure – 06
7.	General Purchase Conditions	Annexure – 07
8.	Statement of Deviations	Annexure – 08
9.	Guidelines for Submission of Bank Guarantee	Annexure – 09
10.	Proforma for Performance Bank Guarantee	Annexure – 10
11.	Technical Bid Form	Annexure – 11
12.	. Commercial Bid form	Annexure - 12

Annexure - 01 Schedule of Quantity

Supply, Installation and Commissioning of Scientific Equipment at Department of Biochemistry, Osmania University.

Sl.No.	Description of Work	Quantity (Nos)
Suppl	y, Installation and Commissioning of	
1.	Fluorescent Microscope inverted	1
2.	Cold High Speed Centrifuge with rotors	1
3.	Flow cytometer	1
4.	Multimode Plate Reader	1
5.	Zeta Potentiometer	1
6.	Fabrication of Tissue Culture room with pressure Regulators	1

Annexure - 02

ELIGIBILITY CRITERIA AND SPECIAL TERMS AND CONDITIONS

A. ELIGIBILITY CRITERIA FOR TENDERS:

- 1. The Company / the tenderer should be in existence for the last 5 Years
- The Company/ the tenderer should have earlier supplied atleast TWO or THREE TIMES the quantities of the items being in the current tender in any of the last two financial years to the Universities or Research organizations.
- 3. The Company or tenderer should have at least one Service Center in Telangana.
- 4. The tenderer should be a Manufacturer or the authorized representative of equipment or other respective products/items.
- 5. The tenderer/tenderer's group should have a turnover a minimum of more than 1 crore in each of the last two financial years.
- 6. Retunes filled for last two years.

B. SPECIAL TERMS AND CONDITIONS:

- 1. **Delivery Period**: The delivery should be made within 50 days from the date of receipt of purchase order by the tenderer.
- 2. Warranty: All items supplied by the tenderer shall be guaranteed against any defects and the tenderer should provide time-to-time operational maintenance support for a period of 3 years (On Site comprehensive Warranty). The said warranty should cover all Hardware products & Software. The warranty and service shall be provided directly from the manufacturer. The supplied products shall reflect with product codes/part number for three year warranty on their website/or on the call centers. Necessary guarantee certificates shall accompany the supplies. The tenderer shall be liable to rectify any defects that may be found in the equipment supplied free of cost.

- 3. **Installation**: The installation should be done at the Dept. of Biochemistry as mentioned in the Annexure 04 enclosed at no extra cost.
- 4. **Response Time:** The response time of the tenderer to attend to any complaint upon receipt of the complaint/information from the user should not be more than 24 hours.

TENDERER

SEAL

Annexure - 03

Documents to be enclosed

The following documents should be submitted along with the bidding form otherwise the bidders are liable to be disqualified.

- 1. All the Supporting documents in respect of Eligibility Criteria. i.e
 - Registration/Incorporation Certificate in support of the existence of the company for required number of
 years as per the tender schedule eligibility criteria.
 - Copies of Purchase Orders in support of the supply of required quantities of items in any of the last two financial years as per the tender schedule eligibility criteria.
 - List mentioning the addresses and contact persons with phone numbers of the Service Centers present through out Telangana.
 - Supporting Documents indicating that the tenderer is the Manufacturer or the OEM Authorized representative of the corresponding items/products.
 - Audited financial statements that are Balance Sheet, Income & Expenditure and Profit & Loss accounts in support of having required turnover in each of the last two financial years.
- Annexure 02 duly signed & office Seal affixed as a token of acceptance of Special Terms and Conditions.
- 3. Documents in support of Technical Specifications for the scientific equipment as mentioned in Annexure 05 clearly mentioning the make and model.

Supporting Documents if any for equipment.

- 4. The latest Income Tax Clearance Certificate and IT returns for the last two financial years
- 5. The list of customers, to whom the bidder had supplied identical materials in the past along with P.O. details and performance report, if any.
- 6. Annexure 08

Note:

- 1. The Coordinator, DRS-III Programme, Dept. of Biochemistry, O.U will not hold any risk and responsibility for non visibility of the scanned document or non receipt of hardcopies or loss in transit.
- 2. The Documents that are received in time will only be considered for Technical Bid Evaluation.
- 3. The tenderer will be disqualified at any stage of the tender process, if found to have misled or furnished false information in the forms/Statements/Certificates submitted in proof of 1 to 5 above.

Address where the equipment to be installed

Coordinator DRS-III PROGRAMME DEPARTMENT OF BIOCHEMISTRY OSMANIA UNIVERSITY, HYDERABAD-500 007, T.S., INDIA

Annexure - 05

TECHNICAL SPECIFICATIONS

Schedule: Scientific Equipment

1	Fluorescent Microscope Inverted	Complied/
		Not Complied
1.	Binocular head, Eyepiece: High point, Plain stage, Glass insert,	
	Attachable mechanical stage, X-Y Coaxial control, moving	
	range, Auxiliary stages, Up to >5Mega Pixel, LCD screen,	
	Storage: SD card slot with memory card (4G capacity)	
2.	Filter: UV/Blue/Green	
3.	Red excitation: BP 570-650	
4.	Blue excitation: BP460-490	
5.	Green excitation: BP480-550	
6.	Essential Spares have to be offered.	
7.	Service contract for 3 years	
8.	Magnification 10X,20X,40X	
9.	Training on instrument to our persons (on site / at your site)	
10.	Quote price for CIP Hyderabad. Instruments clearance	
	Certificate (DSIR) will be provided by Osmania University	
11.	Warranty period (in Years)	
12.	Additional information / concessions	

2. COLD HIGH SPEED CENTRIFUGE WITH ROTORS	Complied/ Not Complied
Specifications:	
1. Maximum Speed: 21,000 rpm	
2. Max. Force: 51,070 x g	
3. Max Rotor Capacity: 3 liters	

4. Braking system: Dynamic, electric brake	
5. Temperature Set Range: - 20°C to 40°C	
6. Temperature Control Range: +2°C to 37°C at maximum rotor speed	
7. Temperature Control Accuracy: ±1°C of set temperature (after stabilization)	
8. Speed Control Range: 500 rpm to 21,000 rpm	
Rotor	
9. Rotor to hold 24 - 1.5/2 ml tubes (microfuge rotor)	
10. Rotor to holds 50 ml tubes X 6 or 8	
Additional Information:	
11. Service contract for 3 years.	
12. Training on instrument to our persons (on site / at your site)	
13. Quote price for CIP Hyderabad. Instruments clearance Certificate	
(DSIR) will be provided by Osmania University.	
14. Voltage stabilizer	
15. warranty period (in years)	
16. Additional information / concessions	

3. FL(OW CYTOMETER:	Complied/
		Not Complied
1.The system should be capable of detecting at least 5 parameters- FSC, SSC and upto 3 fluorescence colors. It should be equipped with following filters- Green-B (525/30 nm), Yellow- B (583/26 nm), Red-B (695/50 nm).		
2.	System should be equipped with 1 lasers-488 nm.	
3.	The instrument shall be robust so that it does not need laser alignment adjustment by the user.	
4.	The fluidics design of the system should be such so as to significantly minimize bio hazardous waste.	
5.	The System should be able to perform Absolute volumetric Direct Cell count assays without use of Any Reference Beads.	
6.	The instrument shall be able to detect particle size in the range of 0.4-60µm.	
7.	autosampler to use 96 well plate.	
8.	The instrument should allow 1.2ml to 1.5ml microtubes and high throughput 96 well plate sample format.	
	9	<u>I</u>

9. The System should have Application software which should perform both the data acquisition and Analysis. Software should also include Assay Suite for all Turnkey assays available with the system as well as user defined analysis.	
10. The system should be provided with Dell™ laptop running Windows® 7 Ultimate (32-bit only), and including Microsoft Excel. Minimum configuration: Intel Core i5-3210M processor (2.5GHz, 3M cache); 4 GB, DDR3-1600 MHz SDRAM; 320 GB hard drive; 2 USB portsThe system should be compact bench top Flowcytometer for routine assays including Instrument control, Real time acquisition, data analysis & representation in single/dual parameter histogramor dual parameter plots during acquisition.	
11. Additional information / concessions	

1. The reader should be capable of measuring Absorbance (UV and Visible), Fluorescence Intensity Top, Fluorescence intensity Bottom reading TRF, FRET, Luminescence Glow, Luminescence Flash, and Quartz based DNA/ RNA Quantification plate with sample capacity of 12 Channel in Fluorescence mode and future upgaradation for CO2 and O2 device for cell based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.	4. M	ultimode Plate Reader	Complied/
Visible), Fluorescence Intensity Top, Fluorescence intensity Bottom reading TRF, FRET, Luminescence Glow, Luminescence Flash, and Quartz based DNA/ RNA Quantification plate with sample capacity of 12 Channel in Fluorescence mode and future upgaradation for CO2 and O2 device for cell based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.			Not Complied
TRF, FRET, Luminescence Glow, Luminescence Flash, and Quartz based DNA/ RNA Quantification plate with sample capacity of 12 Channel in Fluorescence mode and future upgaradation for CO2 and O2 device for cell based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Measurement range should be more than 4 OD v) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.		1	
DNA/ RNA Quantification plate with sample capacity of 12 Channel in Fluorescence mode and future upgaradation for CO2 and O2 device for cell based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.	/ *	, 1, , , , , , , , , , , , , , , , , ,	
Fluorescence mode and future upgaradation for CO2 and O2 device for cell based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vi) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.			
based assays. 2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.			
2. The Reader should have Dual Double Monochromatic Technology i.e, 2 monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.		and future upgaradation for CO2 and O2 device for cen	
monochromatic across Excitation and 2 monochromatic across Emission to select any wavelength from UV-NIR range. 3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.	-	r should have Dual Double Monochromatic Technology i.e, 2	
3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.			
3. The Reader should be able to Perform following parameters in absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.	any wavelength	from UV-NIR range.	
absorbance mode: ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: o Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading o Reader should be capable of reading TRF in standard mode.	, ,		
ii) Wavelength Range from 250 nm to 1000nm with 1nm Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.		9 .	
Increment iii) Wavelength Selection should be monochromatic iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.			
iv) Wavelength Measurement range should be more than 4 OD v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.	11)		
v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: • Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading • Reader should be capable of reading TRF in standard mode.	iii)	Wavelength Selection should be monochromatic	
v) Wavelength Accuracy should be less than 1nm vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: • Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading • Reader should be capable of reading TRF in standard mode.	iv)		
vi) Wavelength Reproducibility should be less than 1nm vii) Detector- UV Silicon Photodiode. 4. The Reader should be able to perform following parameters in Fluorescence mode: • Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading • Reader should be capable of reading TRF in standard mode.	v)		
4. The Reader should be able to perform following parameters in Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading O Reader should be capable of reading TRF in standard mode.	· · · · · · · · · · · · · · · · · · ·		
Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading O Reader should be capable of reading TRF in standard mode.	· · · · · · · · · · · · · · · · · · ·		
Fluorescence mode: O Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading O Reader should be capable of reading TRF in standard mode.			
 Wavelength Range from 250 nm to 900nm across Excitation and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode. 		• 9 •	
and 300 nm to 900nm across Emission for Fluorescence Top reading and Bottom Reading Reader should be capable of reading TRF in standard mode.			
Reader should be capable of reading TRF in standard mode.			
		· ·	
O Wavelength Selection should be monochromatic for			
Fluorescence top and bottom reading.			
o Fluorescence Top reading sensitivity should be less than			
200amol/well			
o Fluorescence Bottom reading sensitivity should be less than 1.5fmol/well		· · · · · · · · · · · · · · · · · · ·	

o TRF sensitivity should be less than 100amol/well				
0	Detector – Sensitive PMT			
0	Wavelength Accuracy should be less than 2nm			
0	Wavelength Reproducibility should be less than 1nm			
0	Fluorescence Based quantification Plate for RNA/DNA application.			
5 The reader	should be able to perform following parameters in Luminescence			
mode:	graduate to perform renorming parameters in La nimicoconics			
mode.				
0	Wavelength Range from 350 nm to 650nm with 1nm Increment			
0	Reader should perform dual color luminescence.			
0	Dynamic range should be more than 7 order of Magnitude.			
0	Luminescence Sensitivity should be less than 10 pM for 384 well			
	plates.			
0	Luminescence ATP detection limit should be less than 3 fmol/well			
0	Detector- Photon counting low dark PMT.			
0	Injector volume should be minimum 5ul and maximum 1ml.			
0	Dead volume of 100ul.			
0	Reader must be able read luminescence Glow.			
6.The reader s	hould be able to perform following parameters in general mode:			
0	Plate shaking: Linear and orbital.			
0	Light Source: UV Xenon Flash Lamp.			
 Temperature Control: Ambient +5 Deg C – 42 Dec C Plate forma: 6 to 384 well plate, Cuvette Port, Quantification 				
0				
	plate.			
0	Gas Control module with 02 and C02 (Upgradeable)			
7. Essen				
0	The system should be supplied along with suitable PC			
0	The software should have function for drag and drop for assay			
	sequence and data reduction which provides an automatic export of measurement parameters into result files in a user specified			
Additional Information:				
8. Service contract for 3 years.				
9. Training on instrument to our persons (on site / at your site)				
10. Quote price for CIP Hyderabad. Instruments clearance Certificate				
(DSIR) will be provided by Osmania University.				
11. Voltage stabilizer				
	12. warranty period (in years)			
13. Additional information / concessions				

5. Specifications for High Concentration Nano Size and Zeta Potential Measurement system:

Complied/
Not Complied

The Dynamic Light Scattering based system should be capable of measuring hydrodynamic properties like diameter, correlation function, sample polydispersity, zeta potential, conductance, aggregation propensity, approximate molecular weight of materials with dimension in nanoscale and sub-micron to micron scale.

The system should be capable for measurement of highly diluted to very concentrated samples.

System Features:

- Dynamic Light Scattering for particle size and Phase Analysis Light Scattering for electrophoretic mobility, zeta potential and charge.
- Particle Size measurement range: ≤ 0.5 nm to ≥ 5 microns.
- Should have customizable dust limit annotation in software.
- Concentration capability to measure particle Size and zeta potential: minimum 0.1 mg/ml of weakly scattering material such as protein for size or less to maximum 40% w/v for zeta potential or better
- Zeta Potential measurement range: -500mV to +500mV or better
- Zeta Potential measurement Technique: PALS/M3-PALS or equivalent advanced signal-processing techniques for fast reproducible data generation.
- Temperature control: Internal sample temperature control/peltier facility should be available with minimum temperature $\leq 2^0$ C to 90^0 C or higher to cover the measurement over a large range of biological activity occurring at various temperature domain.
- Conductivity Range for Mobility measurements: system should be able to perform measurements at high salt concentrations.
- Optional to upgrade to do advanced studies such as DLS-microrheology, surface zeta potential studies would be preferable.

Optics and detectors:

- Laser Power: 4mW or higher power with auto attenuation feature for higher sensitivity of the instrument.
- Measurement Angle: Measurement angles should include preferable non-invasive backscattering (160° or higher) for determining size and forward scattering (15° or lower) for determining mobility/zeta potential.
- Availability of position variation within cell (vendors should specify number of positions) to collect scattered light so as to minimise multiple scattering effects.
- Detector: High sensitivity Avalanche Photodiode Detection (APD) for DLS detection and mobility measurements.

Sample Cells:

- Quartz cuvettes, Flow cells, Dip cells, Disposable cuvettes for size and zeta potential measurements of aqueous and organic samples should be supplied with the instrument as standard.
- For particle size: cuvettes with minimum sample volume about 50 microlitres or less.
- For zeta potential: cuvettes with minimum sample volume about 100 microlitres or less preferably 20ul.
- Flow cell: should work at about 1ml/minute or less, non-aqueous solvent compatible. The same cell should be suitable for both manual and automated measurements.
- Sufficient sample cuvettes (both disposable and quartz) for future applications.
- Suitable standard solutions for the calibration of all the detection modes (DLS, Zeta Potential, SLS) should be provided with the system
- Nano smaple measurment plate for DNA and protein estimations

Software:

- Sutable lap top with Standard research grade software should include capability to measure and report size, zeta potential, molecular weight and their distribution along with Trend Analysis methods.
- It must be pre-loaded with Dispersion Liquid Properties, editable diluents' list with automatic parameters determination including the viscosity library.
- It should use standard algorithms such as NNLS, CONTIN, regularization and Cumulant etc. for sizing and various models such as Scmoulski, Huckel, White-Mangelsdorf for zeta potential.
- The software should have facility to export measured data and result into ASCII format, to facilitate further data processing.
- User defined and lockable multi-component standard operating procedures (SOP).
- Feature for molecular weight measurements.
- Feature for determination of particle charge via simultaneous analysis of hydrodynamic radius and electrophoretic mobility, and determination of the isoelectric point of a sample via interpolation of charge vs pH.
- The software should have capabilities to carry out protein studies and its state such as melting point study/crystal screening.

Warranty:

Comprehensive onsite warranty for **One** year from the date of installation.

6.Fabrication of Tissue Culture room with pressure Regulators	Complied/
	Not Complied
Clean air room fabrication (size of the room 20 x 12 ft) for positive pressure	
(for cell culture work station) with air conditioner (2ton)	
Note: Please visit and inspect the room dimensions in Dept. of Biochemistry,	
Osmania University.	

'Instructions to Bidders'

Due date of submission of bids : 03-08-2017 at 03.00 P.M

Due date of opening of Technical Bids : 03-08-2017 at 04.00 P.M

Due date of opening of Price Bids : 03-08-2017 at 04.00 P.M

The instructions given below must be read very carefully, as failure in compliance with any of these may render the offer liable for rejection. If a bidder has any doubt about the meaning of any stipulation herein, General Purchase Condition, specification of materials or any other enclosed document, he should immediately obtain the clarification/information in writing from the Hyderabad.

1. Earnest Money Deposit (EMD)

- 1.1. Offers must accompany EMD by way of DD/Pay Order, drawn in favor of "Coordinator, DRS-III Programme, Department of Biochemistry, Osmania University, Hyderabad, Telangana", and payable at any of the scheduled bank branches situated in Hyderabad. Submission of EMD in any other form shall not be considered and treated as disqualification. The DD [EMD] should be kept separately in a sealed cover duly super scribing 'EMD'.
- 1.2. The EMD shall be forfeited, if
 - i) Bidder withdraws the bid before expiry of its validity.
 - ii) Successful bidder does not accept the order or fails to enter into a contract within validity period of offer.
 - iii) Successful bidder fails to furnish Security Deposit within one month from the date of issue of P.O.
- 1.3. The EMD of unsuccessful bidders shall be returned as soon as the tenders are finalized. Transaction fee payable to C1 India is non refundable.

2. Acceptance of University Payment Terms The standard payment terms shall be

100% of all inclusive price of the materials will be paid on receipt and acceptance of the material in good condition, installation and commissioning of the equipment supplied at site and after furnishing of Bank Guarantee for 10% of the contract covering the guarantee period subject to penal provisions applicable in case of delay in supply and commissioning as per the condition 2.4 of annexure-07.

3 **Price Basis**

The prices quoted shall be on FOR destination basis as per annexure-04, and also shall include:

- (a) Transit Insurance:
- (b) All Taxes, Duties and Levies.

4. Cost Compensation for Deviation:

Deviations specifically declared by the bidders in respective Deviation Schedules of Bid Proposal Sheets only will be taken into account for the purpose of evaluation. The bidders are required to declare the prices for the withdrawal of the deviations declared by them in the Deviation Schedules.

Offers should strictly be in conformity with specifications/ drawings/ samples as stipulated in the enquiry. In case no deviations are indicated, it shall be taken for granted that item (s) has/ have been offered strictly as per the requirements given in the enquiry.

5.0 **Delivery**

5.1 Delivery Schedule

The Supply, Installation, and Commissioning of the material: The bidder shall offer his best, realistic and firm delivery, which shall be specific and guaranteed. Delivery period shall be reckoned from the date of P.O. which is the first intimation of acceptance of bidder's offer. Final date of delivery shall be evidenced by date of dispatch of materials as per transporter's Lorry Receipt/ Goods Receipt/ RR/ PWB/ AWB. For delivery beyond contractual delivery period, provisions of 'General Purchase Conditions' shall apply.

5.2. Early Delivery:

It shall be noted that if an order is placed on higher bidder, in preference to lowest acceptable offer, in consideration of an earlier delivery, the bidder shall be liable to the University, the difference between the ordered rate (s) and the rate (s) quoted by the lowest acceptable bid in case the tenderer fails to complete the supply in terms of such order within the date (s) of delivery specified in the tender and incorporated in the order. This is without prejudice to other rights of the University under terms of order.

- 6. It is not binding on University to accept the lowest or any bid. The University reserves the right to place orders for individual items with different bidders and to revise the quantities at the time of placing the order and in such event also, the quoted rates, terms and conditions shall apply. The order for the materials may also be split up between different bidders to facilitate quick delivery of critically required materials. University further reserves the right to accept or reject any/ all bids without assigning any reason thereof. Final decision on bids also depends on the components/accessories/additional features offered
- 7. The University at its sole discretion unilaterally may change the quantities to the extent of $\pm 30\%$ as indicated in tender enquiry. The successful bidder shall be bound to supply these quantities at the same rate and on the same terms and conditions.
- 8. Tenderers shall fill in the enclosed proforma with regard to deviations / variations (Annexure-08) and submit the same along with their offer.
- 9. Orders placed against this tender enquiry shall be subject to 'General Purchase Conditions' of University, a copy of which is enclosed. Bidders are requested to confirm acceptance of these conditions in to in their offer.

10. Make/Brand of items offered shall be specified failing which offers are liable to be rejected. It shall be appreciated if one copy of detailed descriptive literature / pamphlets is enclosed along with the offer which may help technical evaluation. In case material offered is ISI marked/tested at any Govt. recognized test house; copies of relevant certificates shall be furnished along with the offer.

Annexure – 07

GENERAL PURCHASE CONDITIONS

SECTION- I GENERAL

1.1 The following terms and expressions used herein shall have the meaning as indicated therein: Supplier / Vendors: shall mean the individual firm or company whether incorporated or otherwise in whose name the purchase order is addressed and shall include its permitted assignees and successors. Purchaser: shall mean Coordinator, Dept. of Biochemistry, Hyderabad.

1.2. Reference:

The purchase order number must appear on all the correspondence, packing slips, invoices, drawings or any other document or paper connected with the purchase order.

1.3. Waiver:

Any waiver by the purchaser of the terms and conditions of the purchase order shall not constitute any right for subsequent waiver of any other terms or conditions.

1.4 Sub-letting and Assignment:

The supplier shall not, save with prior consent in writing of the University, sublet, transfer or assign this order or any part thereof or interest therein or benefit or advantage thereof in any manner, whatsoever provided nevertheless that any such consent shall not relieve the supplier from any obligation, duty or responsibilities under the contract.

1.5. Information Provided by the University:

All drawings, data and documentation that are given to the supplier by the University for the execution of the order are the property of the University and shall be returned when demanded. Except for the purpose of executing the order of the University, the supplier shall ensure that the above documents are not used for any other purpose. The supplier shall further ensure that the information given by the University is not disclosed to any person, firm, body, corporate and/or authority and make every effort to keep the above information strictly confidential. All such information shall remain the absolute property of the University.

1.6. Supplier's Liability:

Supplier hereby accepts full responsibility and indemnifies the University and shall hold the University harmless from all acts of omission and commission on the part of the supplier, his agents, his subcontractors and employees in execution of the purchase order. The supplier also agrees to defend and hereby undertakes to indemnify the University and also hold it harmless from any and all claims for injury to or death of any and all persons including but not limited to his/her employees and for damage to the property arising out of or in connection with the performance of the work under the purchase order.

1.7. Access to supplier's Premises:

The University and / or its authorized representative shall be provided access to the supplier's and / or his sub-contractor's premises, at any time during the pendency of the purchase order, for expediting the supplies, inspection, checking etc.

1.8. Modifications

The purchase order constitutes an entire agreement between the parties hereto. Any modification to this order shall become binding only upon the same being confirmed in writing duly signed by both the parties.

1.9 Inspection/checking/testing

All materials/ equipment to be supplied against this purchase order shall be subject to inspection/ checking/ testing by the University or its authorized representative at all stages and places, before, during and after the manufacture. All these tests shall be carried out in the presence of authorized representative of the University. Supplier shall notify the University for inspection of materials/ equipment when they are ready, giving atleast 10 days notice. If upon receipt at our Stores, the material/equipment does not meet the specifications, they shall be rejected and returned to the supplier for repair/modification etc. or for replacement. In such cases all expenses including to- and- fro freight, re-packing charges, transit insurance etc shall be to the account of supplier.

Inspection by the authorized representative of the University or failure of the University to inspect the material/equipment shall not relieve the supplier of any responsibility or liability under this purchase order in respect of such material/ equipment and it shall not be interpreted in anyway to imply acceptance thereof by the University. Whenever specifically asked for by the University, the supplier shall arrange for inspection/testing by Institutional Agencies such as Lloyds Register of Industrial Services, Boiler Inspectorate, RITES. In such cases supplier shall adhere to the inspection/ testing procedures laid down by such agencies. All expenses in this regard including inspection fees shall be to the suppliers account unless agreed to the contrary and specified in the purchase order.

1.10 Packing and Marking:

All materials/equipment shall be securely packed to the requirements of transportation by Air/Rail/ Road/ Sea. All exposed services/ connections/, protrusions shall be properly protected. All unexposed part shall be packed with due care and the packages should bear the words' handle with care'. The packing requirements of Air/Rail/ Road transport shall be complied with so as to obtain clear Airway/Railway Receipt/ Lorry Receipt i.e. without any qualifying remarks.

All packages and unpacked materials shall be marked with the name of Consignor, Consignee, Purchase order No., gross and Net weight, sign of handling, if any, with indelible paint in English atleast at two places. In case of bundles, metallic plates marked with the above details shall be tagged to such bundles.

1.11 Despatch of Materials:

The supplier is responsible for the safe delivery of the goods in good condition at destination stores. The supplier should acquaint himself of the conditions relating to handling and transport of the goods to destination and shall include and provide for security and protective packing of the goods so as to avoid damage in transit.

1.12 Validity of offers:

The offers shall be valid for a **period of 180 days** (**Depending on the type of equipment**) from **the date of opening of bids**. The period of validity cannot be counted from any other date other than the date of opening the bids. During this period the tenderer shall not be permitted to withdraw or vary his offer made and if the tenderer does so, **the EMD shall be forfeited.**

1.13 Jurisdiction:

All and any disputes or difference arising out of or touching this order shall be decided only by the Courts or Tribunals situated in Hyderabad.

SECTION - II

FINANCIAL

2.1. **Prices:**

Prices quoted shall be inclusive of all taxes and firm till completion of the programme.

2.2. Terms of Payment:

- 2.2.1. Payments by the purchaser shall be made through Account Payee cheques only. **Bank** charges if any shall be borne by the supplier.
- 2.2.2. If the supplier has received any overpayments by mistake or if any amounts are due to the University from the supplier due to any other reasons and when it is not possible to recover such amount under the present purchase order, the University reserves the right to collect the same from any other amounts and/ or Bank Guarantee given by the supplier due to or with the University.

2.3 Liquidated Damages/ Failure and Termination:

- 2.3.1 In the event of any delay in the supply of material beyond the stipulated date of completion including any extension permitted in writing, the University reserves the right to recover from the supplier a sum equivalent to 0.5% of the value of delayed materials/equipment for each week of delay and part thereof subject to a maximum of 5% of the total value of the order.
- 2.3.2 Alternatively, the University reserves the right to give the contract elsewhere at the sole risk and cost of the supplier and recover all such extra cost incurred by the University in procuring the materials from the other source.
- 2.3.3 Alternatively University may cancel the Purchase Order completely or partly without prejudice to its right under the alternatives mentioned above.
- 2.3.4 In case of recourse to alternative 2.4.2 and 2.4.3. above, the University shall have the right to repurchase the materials which is readily available in the market to meet the urgency requirements caused by supplier's failure to comply with the scheduled delivery period irrespective of the fact whether the material/ equipment is similar or not.

2.4 **Delivery Schedule:**

Time is essence of this order and no delay shall be allowed in the delivery time/ delivery schedule mentioned in the purchase order.

2.5 **Performance Bank Guarantee:**

The supplier shall ensure that all materials/equipment under this purchase order shall conform to University's requirements and specifications. An additional security in the form of Performance Bank Guarantee is essential for satisfactory performance of the equipment over a period of time. In view of this, the supplier shall be required to furnish a Bank Guarantee (10% of order value)

as follows against any manufacturing defects/ poor workmanship/poor performance. In case any deficiencies are found during this period, the same shall be repaired/rectified/replaced free of cost. BG shall be from any Scheduled Bank or any other bank as approved by University from time to time in the prescribed Performa.

a. Bank Guarantee for 10% of the order value with validity up to warranty period from the date of installation of equipment.

The University shall at its discretion have recourse to the said bank guarantee for recovery of any or all amount due from, the supplier in connection with the purchase order including of guarantee obligations. Checking/approval of supplier's drawings, inspection and acceptance of materials/equipment furnishing to effect shipment and/or work done by erection, installation and commissioning of the equipment by University or any other agency on behalf of the University shall not in way relieve the supplier from the responsibility for proper performance during the guarantee period.

2.6 **Insurance:**

Supplier shall arrange suitable insurance cover at his risk and cost.

2.7 Removal of rejected goods and Replacement:

- a) If upon delivery, the material/equipment is found not in conformity with the specifications stipulated in the purchase order, whether inspected and approved earlier or otherwise, such material / equipment will be rejected by the University or his authorized representative. A notice to this effect shall be issued to the supplier, normally within 30 days from the date of receipt of materials at our Stores.
- b) Supplier shall arrange suitable replacement supplies and remove the rejected goods within 30 days from the date of notice failing which, the goods shall be dispatched to vendor by road transport on 'freight to pay basis' at supplier's risk and cost.
- c) External damages or shortages that are prima-facie as a result of rough handling in transit or due to defective packing shall be intimated to the supplier within a period of one month of the receipt of the materials, spares etc. In case of internal defects, damages or shortages of any internal parts, which cannot ordinarily be detected on a superficial visual examination, due to bad handling in transit or defective packing, or any other reason, it should be intimated to the supplier within 3 months from the date of receipt of the material. In either case the damaged or defective materials should be replaced by the supplier free of cost.
- d) If no steps are taken within 15 days of receipt of intimation of defects or such other reasonable time as the University may deem proper to afford, the University may without prejudice to its other rights and remedies arrange for repairs/rectification of the defective materials or replace them entirely and recover the expenditure incurred on account of these actions from the deposits such as EMD, SD and performance guarantees or other monies available with the University or by resorting to legal action.

2.8 Force Majeure:

- 2.8.1 The supplier shall not be liable for delay or failing to supply the material for reasons of Force Majeure such as Act of God, Act of War, Act of Public Enemy, Natural calamities, Fires, Floods, Frost, Strikes. Lockouts etc. Only those causes which have duration of more than 7 days shall be considered for force majeure.
- 2.8.2 The supplier shall within 10 days from the beginning of such delay notify the University in writing the cause of delay. The University shall verify the facts and grant such extension of time as facts justify.

- 2.8.3 No price variation shall be allowed during the period of force majeure and liquidated damages would not be levied for this period.
- 2.8.4 At the option of University, the order may be cancelled. Such cancellation, would be without any liability whatsoever on the part of the University. In the event of such cancellation, supplier shall refund any amount advanced or paid to him by the University and deliver back any materials issued to him by the University and release facilities, if any provided by the University.

DEVIATIONS

Bidder's Name and Address

To Registra Hyderab		nania Univ	ersity							
Dear Sir	rs, Sub:-	Supply o	f	dated*	**	again 	ıst enqui	ry		
Nooffer. The be executed as the second s	ations v nese de uted as nd in o	and tend which are in viations and per specification of the pur offer of	er document irrespective of and variations arcations and	are the only its for the all of what-so-eve are exhaustive tender docume se stated below	oove ment r has been s e. Except fo ents. Furthe	ioned subject stated to the c r these deviat er, we agree the	et supplications, the hose add	ies again anywhere entire su litional co	else of pplies ondition	quiry of our shall ons, if
SL.No.		cription of	Deviation	Ref. of pag	e, Clause	*Monetary	•			the val
						Rs.	(in figur	res) Rs. (in wor	rds)
conditio	ns/devi	iations and	e extra (i.e.	: Here the ter in addition to the condition ecessary).	o the rates	quoted by	him) for	r withdra	wal o	f his
Signatur	re:									
Designa	tion:									

Guide-lines for Submission of Bank Guarantee towards Performance Security

The Bank Guarantee shall fulfill the following conditions failing which it shall not be considered valid:

- 1. Bank Guarantee shall be executed on non-judicial stamp paper of applicable value purchased in the name of bank.
- 2. Non-judicial stamp paper shall be used within 6 months from the date of purchase. Bank Guarantee executed on the stamp paper of more than 6 months old shall be treated as invalid.
- 3. The contents of the Bank Guarantee shall be as per our proforma (Annexure- 10).
- 4. The Bank Guarantee should be executed by a scheduled bank or banks viz.
- 5. The executor of Bank Guarantee (Bank Authority) should mention the Power of Attorney No. and date executed in his/her favour authorizing him/her to sign the document or produce the Photostat copy of Power of Attorney.
- 6. All conditions, corrections, deletion in the Bank Guarantee should be authenticated by signature of Bank Officials signing the Bank Guarantee.
- 7. Each page of Bank Guarantee shall bear signature and seal of the Bank.
- 8. Two persons should sign as witnesses mentioning their full name and address.

Coordinator
DRS-III Programme
Dept. of Biochemistry, OU

BANK GUARANTEE PROFORMA FOR PERFORMANCE SECURITY

This	agree	ment l	nas to be	executed	on a Non-Ju	dicial sta	amped pap	er worth R	s.100/-			
Whe	reas	the	• • • • • • • • • • • • • • • • • • • •		. here-in-a	fter ca	lled (The	e Bidder)	has su	ıbmitted	their	bid
dated	d			for	the supply	of			(Here-	in-after	called	"the
Bid"	")				KNOW	ALL	MEN	l by	these	prese	nts	that
we									(Hereir	nafter c	alled	"the
Bank	k") are	e bour	nd unto I	Registrar,	Osmania Un	iversity	, Hyderab	ad, Telanç	jana. He	reinafter	called	."the
purc	haser'	') in th	e sum o	f Rs				for	which pa	yment wi	ill and	truly
to be	e mad	le to t	he said	purchaser,	the bank b	inds itse	elf, its suc	ecessors an	d assign	s by thes	se pres	sents.
Seal	ed w	ith th	e com	non Seal	of the sa	id Ban	k this					day
of				200								
1. 2.	When letter. When	n the /letter n the s	successfor of inden	ul tendere t/purchase ıl tenderer	er does not order. fails to furneptance let	accept	the order security of	leposit with	hin 30 da	ys from	the da	ite of
3. 4.	When We us writted supplied legality by it occur.	n tende inderta en den ier and ity of t is due rred co	erer alter ake to pa nand withou the dema to it ow ondition of	s his price by to the p shout the p at question and provide ying to any or condition	or reasons ou s or withdraw burchaser the burchaser ha ling the right ed that in its of the occur ons.	ws his of e above ving to t of Units demand arrence of	amount we substantial versity to dishe above the above pring our li	the validity within one was the his demake such haser will we mention ability und	week upo and, with h demand note that led condi-	on receipt nout refer d or the p the amou tions, spe	rring to proprie unt cla ecifyin	to the ety or imed g the
	rights liabil	unde ity the	er this greafter.	uarantee s	hall be forf	eited an	d we sha	ll be relie	ved and	discharge	ed froi	m all
					Bank h the previou							
					Day							
	Signa	iture:		Desi	gnation.		Address	-	Sea	1.		

ANNEXURE-11

TECHNICAL BID FORM

S.No	Documents	Copy submitted or not (Yes/No)
1	The Company / the tenderer should be in existence for the last 5 Years	
	As per Annexure -02	
2	The Company/ the tenderer should have earlier supplied TWO or	
	THREE TIMES the quantities of the items being in the current tender in	
	any of the last two financial years to the Universities or Research	
	organizations.	
3	The tenderer should be a Manufacturer or the authorized representative	
	of equipment or other respective products/items	
4	The tenderer/tenderer's group should have a turnover a minimum of	
	more than 5 times the Bid value of the tenderer in each of the last two	
	financial years.	
5	Audited financial statements, that is Balance Sheet, Income &	
	Expenditure and Profit & Loss accounts in support of having required	
	turnover in each of the last two financial years.	
6	Technical Specifications of Annexure – 05 (Complied or Not-Complied	
	report) along with supporting documents of the items bided (items 1 to	
	9), for the scientific equipment clearly mentioning the make and model	
7	The latest Income Tax Clearance Certificate and IT retunes for the last	
	two financial years	
8	List mentioning the addresses and contact persons with phone numbers	
	of the Service Centers	
9	The list of customers, to whom the bidder had supplied identical	
	materials in the past along with P.O. details and performance report, if	
	any.	
10	Annexure - 08	

ANNEXURE – 12

COMMERCIAL BID FORM

S.No.	Description of Work	Quantity (No's)	Price in Rs / US \$ (inclusive of all Taxes)
1	Fluorescent Microscope inverted	1	
2	Cold High Speed Centrifuge with rotors	1	
3	Flow cytometer	1	
4	Multimode Plate Reader	1	
5	Zeta Potentiometer	1	
6	Fabrication of Tissue Culture room with pressure Regulators	1	